



Metal & Steel Factory (MSF)
Ministry Of Defence , Government Of India
(ISO 9001, ISO/IEC 17025 NABL ACCREDITED)
Ishapore , , Pin- 743144
E-Mail - , Phone -



UNIQUE CERTIFICATE CODE (UCC)
For Fabric for Personal Protective Equipment (PPE) For COVID-19

This Certificate Code is Issued to M/S PRATAP SYNTHETICS LTD. with Registered Address as:

M/S PRATAP SYNTHETICS LTD.
21, BENARAS ROAD, SALKIA
HOWRAH , WEST BENGAL , 711106

Against their Fabric for Personal Protective Equipment (PPE) test sample with sample No - FABRIC 71 GSM submitted to us on 15-May-2020 and tested on 27-May-2020, the sample has passed the Synthetic Blood Penetration Test conducted as per and confirms to the technical requirement of Ministry of Health & Family Welfare guideline dated-02-03-2020. The test result is annexed to this certificate.

The sample's UCC code is: **MSF014590520**
Which is valid upto: 27-Nov-2020

The affidavit furnished by the applicant in relation to their sample is annexed alongwith this certificate.

This certificate is based on the sample submitted by the applicant. No other test/verification has been conducted by the testing agency other than results in the test report. The procurement agency is advised to conduct prior, due diligence as per the procurement policy, including periodical sample tests during the course of supply of the materials.

-sd-
Authorised Signatory



Metal & Steel Factory (MSF)
Ministry Of Defence , Government Of India
(ISO 9001, ISO/IEC 17025 NABL ACCREDITED)
Ishapore , , Pin- 743144
E-Mail - , Phone -



Report No.-01459
UCC Code-MSF014590520
Report Dated- 29-May-2020

To,

M/S PRATAP SYNTHETICS LTD.
21, BENARAS ROAD, SALKIA
HOWRAH, WEST BENGAL , 711106

Dear Sir/ Madam,

This has reference to the Sample No. FABRIC 71 GSM Dt. 15-May-2020 submitted by you for testing vide your letter No. NIL dated- 14-May-2020.

The results pertaining to your sample(s) are enclosed herewith.

As the test report is computer generated, hence does not require signature.

Important: This report is strictly CONFIDENTIAL. Its use, for publicity, arbitration or as evidence against legal disputes is forbidden. Reference of Sample(s) is as given by the party. The enclosed results are related to the samples tested only. The report shall not be reproduced except in full, without the written approval of the laboratory.

Copy to: M/S HLL Lifercare Ltd. through email: jitesh.hll@gmail.com

For kind information and necessary action please.



Metal & Steel Factory (MSF)
Ministry Of Defence , Government Of India
(ISO 9001, ISO/IEC 17025 NABL ACCREDITED)
Ishapore , , Pin- 743144
E-Mail - , Phone -



UCC No. MSF014590520

UCC Dt. 29-May-2020

Test Report

Fabric for Personal Protective Equipment (PPE) for COVID -19

1. Name of Testing Agency : Metal & Steel Factory (MSF)
2. Date/ Period of Testing : 27-May-2020 to 27-May-2020
3. Test Reference No. : NIL
4. Test Reference Dt. : 14-May-2020
5. Test Reference : ASTM 1670 procedure A
6. Manufacture Name : M/S PRATAP SYNTHETICS LTD.
21, BENARAS ROAD, SALKIA
HOWRAH , WEST BENGAL
7. Environmental Conditions :

S.no	Parameter	Specified Value	Observed Value
1	Relative Humidity (%)	50 to 70	57
2	Temperature(deg C)	16 to 26	25

8. Test Results :

S.no	Sample Nomenclature	Result
1	Fabric portion	Pass

Remarks :


- (i) This UCC has been Issued with reference to MOT letter no F. No 8/4/2020- R and D dated 06-04-2020.
Warning - The above report pertains only to the sample submitted by the customer as the samples are not drawn by the testing agency. These results need not be indicative of the results of bulk lots of the articles under consideration. The onus of maintaining the Traceability between the sample submitted for testing and the bulk lot supplied lies with the customer/ supplier. Samples may be drawn from the bulk lots by neutral Organizations and the same shall be tested to ensure their compliance with the respective quality requirements.
- (ii) Instruction to manufacturers to whom this certificate is issued-The procedures enumerated at Para - 3 and 4 of Ministry of Textiles, Govt. of India letter F no 8/2020-R&D dated 06.04.2020 must be strictly complied with regards to use of this certification.
- (iii) Manufacturer is liable to face legal consequence in the event of any misrepresentation of facts/data during the use of this certificate.
- (iv) The value of thickness observed is 0.28 mm (for information only)

***** End of Report *****

Affidavit Form for Testing & Certification of Proto-Type Sample of Fabrics Required to manufacture PPE Body Coveralls for COVID-19.

1. Name of the Manufacturing Company: Pratap Synthetics Ltd.
2. Registered Address: 21, Benaras Road, Salkia, Howrah 711106
3. Name of Head of Company: Harsh Agarwal
4. (a) Contact No: 9831741122
(b) Official email id: harsh@pratapgroup.com
5. (a) Company Registration No: 038273
(b) GSTIN No: 19AABCP4794Q1ZH
6. Name and Address of Manufacturing Unit of fabric required for PPE Body Coveralls:

Pratap Synthetics Ltd.
Sankrail Industrial Park
Dhulagori, Sankrail, Howrah – 711302



7. Description of Type of Fabric:
 - (a) GSM: 71GSM (28GSM Fabric + 15GSM Lamination + 28GSM Fabric)
 - (b) Thickness: 270 Microns
 - (c) Non-Woven
 - (d) Sandwich Extrusion Lamination with mix of PP and LDPE
 - (e) Coated
8. Declaration:

RAJESH KUMAR BAJPAI
NOTARY GOVT. OF INDIA
Reg. No. 13780/18
C.M.M.S' Court
2 & 3 Bankshall Street
Kolkata-700 001

I Harsh Agarwal, being the authorized person in our company hereby certify that the above information is true to best of our knowledge. I also certify that we are fully aware that the PPE body coveralls are related to the safety of the healthcare professionals, and we will ensure that the quality of our manufactured product meet the standards as mentioned in Ministry of Health & Family Welfare guidelines on the subject dated 02.03.2020 and Ministry of Textiles instructions on quality control of PPE coveralls 06.04.2020.

We are in the business of textiles manufacturing under appropriate government approval, and not traders.

We also understand that, if at any point of time our manufactured and supplied products are found to be not meeting the minimum technical requirement, the UCC will be cancelled and relevant legal provisions can be initiated against us by the law enforcing agencies, including criminal prosecution.


Place: KOLKATA

Date: 6/5/20

Attested
RAJESH KUMAR BAJPAI
NOTARY GOVT. OF INDIA
Kolkata, West Bengal
Regn No. 13780/18

Harsh Agarwal
06/5/2020

Signature of Authorised Signatory with seal



06 MAY 2020